



Certificate of Traceability

Manufacturer Certification

This certifies that **Boditech Med Inc.**, using **Immuno-assay on ichroma Reader** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **Reference Material Institute for Clinical Chemistry Standards (ReCCS) ASRL#1**

The system evaluated was:

Instrument: ichroma Reader	Calibrator Lot: 704101
Reagent Lot: AAPCD68, AAPCD68, AAPCD68	Calibrator Assigned Values: 5.5%, 8.5%, 11%, 14%

Date of Certification: June 1, 2019

Certification Expires: June 1, 2020

Daniel Soubes

NGSP Steering Committee Chair

Randie R Little PhD.

NGSP Network Coordinator

Violeta G. Raneva, PhD.

SRL director/ supervisor